



*An ISO 9001:2015 Certified Hospital*



## MOI TEACHING AND REFERRAL HOSPITAL

Telephone: (+254)053-2033471/2/3/4  
Fax: 2061749  
Email: [ceo@mtrh.go.ke](mailto:ceo@mtrh.go.ke) /[directorsofficemtrh@gmail.com](mailto:directorsofficemtrh@gmail.com)

Nandi Road  
P.O Box 3-30100  
Eldoret, Kenya  
14<sup>th</sup> March, 2024

### ADDENDUM-1

**TENDER NO. MTRH/T/78/2023-2024: TENDER FOR SUPPLY AND DELIVERY OF MEDICAL EQUIPMENT ITEM NO 2 CRITICAL CARE PATIENT MONITORS FOR CRITICAL CARE UNIT.**

Pursuant to provision of section 75(1) of the Public Procurement and Asset Disposal Act, 2015 and section 56 of the Public Procurement and Asset Disposal Regulations 2020, we wish to put forth an addendum in response to the above mentioned tender which were advertised vide MyGOV on 5<sup>th</sup> March, 2024.

Management of Moi Teaching & Referral Hospital has deemed it necessary to provide additional information to the Critical Care Monitors. The desired equipment is for adults and not neonates as earlier indicated. Please find amended specifications.

Other terms and conditions of the tender shall remain as provided in the invitation and the bidding documents.

### 2.0 TECHNICAL SPECIFICATIONS OF CRITICAL CARE CARDIAC MONITORS

S/N	GENERAL OPERATIONAL REQUIREMENTS	REPLACEMENT
A	The monitor shall be configured for NICU application. (4 Marks)	The monitor is for Adults CCU monitors application (4 Marks)
	CARDIAC MONITORS	
B	The monitor display size shall be a minimum of 36 cm (14 inches). (2Marks)	The monitor display size should be at least 22-inch (55cm) (2Marks)

**DR. PHILIP K. KIRWA**

**Ag. CHIEF EXECUTIVE OFFICER**

MOI TEACHING AND REFERRAL HOSPITAL  
CEO  
APPROVED

14 MAR 2024

SIGN.....  
P.O. Box 3 - 30100, ELDORET

*All correspondences should be addressed to the Ag. Chief Executive Officer Visit our Website: [www.mtrh.go.ke](http://www.mtrh.go.ke)*

TO BE THE LEADING MULTI-SPECIALITY HOSPITAL FOR HEALTHCARE, TRAINING AND RESEARCH IN AFRICA

	<b>MONITOR DISPLAY CAPABILITIES</b>	
B	The monitor display size shall have a minimum of 36 cm (14 inches). (1 Marks)	The monitor display size should be at least 22-inch (1 Mark)
	<b>ACCESSORIES, DOCUMENTATION AND GENERAL</b>	
F	Reusable Pediatric NIBP cuffs complete with NIBP hose- 10pcs (1 Mark)	Reusable adults NIBP cuffs complete with NIBP hose- 10pcs (1 Mark)

*Dr. M. M. M. M.*